

EU DECLARATION OF CONFORMITY – No CETF020 rev2.0

Manufacturer: Tunstall Healthcare (UK) Ltd
 Whitley Lodge, Whitley Bridge
 Yorkshire, DN14 0HR
 UK
 Tel: +44 1977 661234

This declaration is issued under the sole responsibility of the manufacturer.
 The manufacturer declares under our own responsibility that the following radio product(s):

| Product name: | Product number(s): | Product Description(s): |
|----------------------|--------------------|---------------------------------------|
| Medication Dispenser | 67005/85 | Pivotell Advance Medication Dispenser |

comply with the requirements of the following directive(s)

- 2014/53/EU Radio Equipment Directive (RED)**
- 2011/65/EU Restriction of the use of certain hazardous substances (RoHS)**

The conformity was checked in accordance with the following harmonised EN standard(s):

| | |
|------------------------------|---|
| EN 300-220-3-1: V2.1.1 | Harmonised Standard covering the essential requirements of article 3.2 of Directive 2014/53/EU; Low duty cycle high reliability equipment, social alarms equipment operating on designated frequencies (869,200 MHz to 869,250 MHz) |
| EN301 489-3 V2.1.1 (2019-03) | Electro Magnetic Compatibility (EMC) standard for radio equipment and services; Part 3: Specific conditions for Short Range Devices (SRD) operating on frequencies between 9kHz and 246GHz; Harmonised Standard covering essential requirements of article 3.1(b) of Directive 2014/53/EU |
| EN301 489-1 V2.2.3 (2019-11) | Electromagnetic compatibility (EMC) standard for radio equipment and services; Part 1: Common technical requirements; Harmonised standard for ElectroMagnetic Compatibility |
| EN 62368-1: 2020+ AC11:2020 | Audio/video, information, and communication technology equipment Part 1: Safety – Requirements |

Signed

Date



18/11/23

Position: Chief Technical Officer